Printing date 18.07.2024 Version number 1 Revision: 18.07.2024

SECTION 1: Identification of the substance/mixture and of the company/undertaking

- · 1.1 Product identifier
- · Trade name: Remin Pro
- · Chemical Identification:

This product is a medical device according to Directive 93/42/EEC concerning medical devices or Regulation (EU) 2017/745 (MDR), which is used invasively or in contact with the body. Therefore, it is exempt from the classification and labelling requirements of Regulation (EC) 1272/2008 (AR CLP, Article 1, paragraph 5 d). A safety data sheet is not required by law for this product. This information is provided on a voluntary basis in the form of this Medical Devices Data Sheet.

- · Product category Dental medical device
- · Article category

The information relevant for the application and for the safety of users and patients is defined in the product-specific directions for use. The instructions for use must be observed.

The product should only be applied by a professionally trained dental practitioner.

- · Application of the substance / the mixture Remineralisation paste
- · 1.3 Details of the supplier of the data sheet
- · Manufacturer/Supplier:

VOCŎ GmbH

Anton-Flettner-Str. 1-3

D-27472 Cuxhaven

info@voco.de

+49 (0) 4721-719-0 Mo - Fr / 08:00h - 16:00h

SECTION 2: Hazards identification

- · 2.1 Classification of the substance or mixture
- · Classification according to Regulation (EC) No 1272/2008

This product is a medical device according to Directive 93/42/EEC concerning medical devices or Regulation (EU) 2017/745 (MDR), which is used invasively or in contact with the body. Therefore, it is exempt from the classification and labelling requirements of Regulation (EC) 1272/2008 (AR CLP, Article 1, paragraph 5 d). The classification criteria of Regulation (EC) No 1272/2008 are not directly applicable to this product, but are considered by us as an orienting basis for assessment. The health and environmental hazards have been assessed and classified on the basis of the information known to us on the components and their reaction processes. The following hazards based on the classification criteria of Regulation (EC) No 1272/2008 for humans and the environment cannot be excluded:

The substance is not classified, according to the CLP regulation.

- · 2.3 Other hazards
- · Results of PBT and vPvB assessment
- · **PBT**: Not applicable.
- · vPvB: Not applicable.

SECTION 3: Composition/information on ingredients

· Additional information: Mixture of non-hazardous substances.

SECTION 4: First aid measures

- · 4.1 Description of first aid measures
- · General information: No special measures required.
- · After inhalation: Supply fresh air; consult doctor in case of complaints.
- · After skin contact:

Rinse with warm water.

If skin irritation continues, consult a doctor.

· After eye contact:

Rinse opened eye for several minutes under running water. If symptoms persist, consult a doctor.

· After swallowing:

If symptoms persist consult doctor.

(Contd. on page 2)

Printing date 18.07.2024 Version number 1 Revision: 18.07.2024

Trade name: Remin Pro

Rinse out mouth and then drink plenty of water.

(Contd. of page 1)

- 4.2 Most important symptoms and effects, both acute and delayed No further relevant information available.
- 4.3 In case of contact with mucous membranes during treatment: No special measures required.

SECTION 5: Firefighting measures

- · 5.1 Extinguishing media
- · Suitable extinguishing agents: Use fire extinguishing methods suitable to surrounding conditions.
- 5.2 Special hazards arising from the substance or mixture No further relevant information available.
- · 5.3 Advice for firefighters
- · Protective equipment: No special measures required.

SECTION 6: Accidental release measures

- · 6.1 Personal precautions, protective equipment and emergency procedures Not required.
- · 6.2 Environmental precautions: No special measures required.
- · 6.3 Methods and material for containment and cleaning up: Pick up mechanically.
- · 6.4 Reference to other sections

See Section 7 for information on safe handling.

See Section 8 for information on personal protection equipment.

See Section 13 for disposal information.

SECTION 7: Handling and storage

7.1 Precautions for safe handling

No special precautions are necessary if used correctly.

For use in dental application only.

Observe the instructions for use! This contains the relevant application and safety information for the use of this product.

- · Information about fire and explosion protection: No special measures required.
- · 7.2 Conditions for safe storage, including any incompatibilities
- · Storage:
- · Requirements to be met by storerooms and receptacles: No special requirements.
- · Information about storage in one common storage facility: Not required.
- · Further information about storage conditions:

Please observe the storage instructions on the packaging and in the instructions for use.

SECTION 8: Exposure controls/personal protection

- · 8.1 Control parameters
- Ingredients with limit values that require monitoring at the workplace:

The product does not contain any relevant quantities of materials with critical values that have to be monitored at the workplace.

· Additional information:

Due to the proportions contained in relation to the amount of product during use and the way it is incorporated into the product, monitoring of specific, workplace-related limit values is not applicable.

- · 8.2 Exposure controls
- · Individual protection measures, such as personal protective equipment
- · General protective and hygienic measures:

The usual precautionary measures are to be adhered to when handling chemicals.

In the context of the use of this product, the hygiene standards customary and general in dental practice, such as hand, mouth and eye protection, must be applied.

Printing date 18.07.2024 Version number 1 Revision: 18.07.2024

Trade name: Remin Pro

(Contd. of page 2)

SECTION 9: Physical and chemical properties

· 9.1 Information on basic physical and chemical properties

· General Information

Physical state Colour: Liquid White

· Odour: according to product designation

· Odour threshold: Not determined. · Melting point/freezing point: Undetermined.

· Boiling point or initial boiling point and boiling

range Undetermined.
• Flammability Not applicable.

· Lower and upper explosion limit

Lower: Not determined.
Upper: Not determined.
Flash point: Not applicable.
Decomposition temperature: Not determined.
pH
Not determined.

· Viscosity:

Kinematic viscosityDynamic:Not determined.Not determined.

·Solubility

water: Fully miscible.
 Partition coefficient n-octanol/water (log value)
 Vapour pressure: Not determined.

Density and/or relative density

Density: Not determined.
Relative density Not determined.
Vapour density Not determined.

· 9.2 Other information

· Appearance:

· Form: Pasty
· Important information on protection of health and

environment, and on safety.

· Auto-ignition temperature: Not determined.

• Explosive properties: Product does not present an explosion hazard.

· Change in condition Not applicable.

SECTION 10: Stability and reactivity

- · 10.1 Reactivity No further relevant information available.
- · 10.2 Chemical stability Stable.
- Thermal decomposition / conditions to be avoided: No decomposition if used according to specifications.
- · 10.3 Possibility of hazardous reactions No dangerous reactions known.
- 10.4 Conditions to avoid No further relevant information available.
- · 10.5 Incompatible materials: No further relevant information available.
- 10.6 Hazardous decomposition products: No dangerous decomposition products known.

SECTION 11: Toxicological information

Other information

As an invasive medical device, the product is subject to relevant laws and standards which ensure the safety and performance of the product when used appropriately and in accordance with the instructions for use. Specific toxicological information is not applicable at this point. The instructions for use must be observed.

Printing date 18.07.2024 Version number 1 Revision: 18.07.2024

Trade name: Remin Pro

(Contd. of page 3)

SECTION 12: Ecological information

· General notes

As an invasive medical device, the product is subject to relevant laws and standards that ensure the safety and performance of the product when used properly and in accordance with the instructions for use. Special environmental information is not applicable at this point. The instructions for use must be observed.

SECTION 13: Disposal considerations

- · 13.1 Waste treatment methods
- · Recommendation

Dispose of in accordance with official regulations. For further information, see the instructions for use.

- · Uncleaned packaging:
- **Recommendation:** Disposal must be made according to official regulations.
- · Recommended cleansing agents: Water, if necessary together with cleansing agents.

	ion	
14.1 UN number or ID number ADR, IMDG, IATA	Void	
14.2 UN proper shipping name ADR, IMDG, IATA	Void	
14.3 Transport hazard class(es)		
ADR, ADN, IMDG, IATA Class	Void	
14.4 Packing group ADR, IMDG, IATA	Void	
14.5 Environmental hazards:	Not applicable.	
14.6 Special precautions for user	Not applicable.	
14.7 Maritime transport in bulk according instruments	g to IMO Not applicable.	
UN "Model Regulation":	Void	

SECTION 15: Regulatory information

· 15.1 Safety, health and environmental regulations/legislation specific for the substance or mixture Regulation (EU) 2017/745

Medical Devices Regulation

Directive 93/42/EEC concerning medical devices

· 15.2 Chemical safety assessment:

The health and environmental hazards have been assessed and classified on the basis of the information known to us on the components and their reaction processes.

A Chemical Safety Assessment has not been carried out.

SECTION 16: Other information

This information is based on our present knowledge. However, this shall not constitute a guarantee for any specific product features and shall not establish a legally valid contractual relationship.

The product to which this data sheet is assigned is a medical device according to the EU Medical Devices Regulation EU 2017/745. Invasive medical devices or medical devices in direct body contact are exempted from the classification and labelling requirements of Regulation (EU) 1272/2008 (AR CLP, Article 1,

(Contd. on page 5)

Printing date 18.07.2024 Version number 1 Revision: 18.07.2024

Trade name: Remin Pro

(Contd. of page 4)

paragraph 5 d). The Medical Devices Regulation does not require a safety data sheet for invasive medical devices or medical devices in direct body contact, as the safe use of the product is indicated in the instructions for use and/or the labelling. Nevertheless, we provide a data sheet for medical devices in order to provide as transparent a source as possible for the assessment of hazards to humans and the environment.

- · Department issuing Datasheet: Knowledge Communication Department
- · Contact:

Global headquarter: VOCO GmbH Anton-Flettner-Str. 1-3 D-27472 Cuxhaven info@voco.de +49 (0) 4721-719-0 Mo - Fr / 08:00h - 16:00h

Australien sponsor address: VOCO Australia Pty Ltd Level 19, 133-145 Castlereagh Street Sydney, NSW 2000

Email: info@voco.com

For further contact information, please visit www.voco.dental

- · Version number of previous version: Not applicable.
- · Abbreviations and acronyms:

ADR: Accord relatif au transport international des marchandises dangereuses par route (European Agreement Concerning the International Carriage of Dangerous Goods by Road)

IMDG: International Maritime Code for Dangerous Goods

IATA: International Air Transport Association

GHS: Globally Harmonised System of Classification and Labelling of Chemicals

PBT: Persistent, Bioaccumulative and Toxic vPvB: very Persistent and very Bioaccumulative

- EH